

APPLICATION		
NEXT ASS'Y	USED ON	VAR

PCB Piezotronics Inc. claims proprietary rights in the information disclosed hereon. Neither it nor any reproduction thereof will be disclosed to others without written consent of PCB Piezotronics Inc.

REVISIONS				
ZONE	REV	DESCRIPTION	ECN	DATE
	C	UPDATED DRAWING	15262	4/22/02
	D	UPDATED TO MEET ISO 16063-21 SPECS	24050	4/10/06

MOUNTING STUD TABLE	
MODEL	THREAD
081B20	1/4-28 TO 1/4-28
081A08	10-32 TO 1/4-28
M081B20	1/4-28 TO M6 X 0.75

MODEL 301A10
BACK-TO-BACK
ACCELEROMETER
STANDARD
(SUPPLIED)

1/4 - 28 THREAD
MOUNTING HOLE

.0004 B

1.188 [30.18] HEX

MODEL 002C10
CO-AXIAL CABLE
10 FEET [3.0 METERS]
LONG (SUPPLIED)

MODEL 482A23
POWER UNIT
(SUPPLIED)

MODEL 012A03
OUTPUT CABLE
3 FEET [0.9 METERS]
LONG (SUPPLIED)

SHAKER
(NOT SUPPLIED)

2.) SUPPLIED WITH MODEL 488A04, EXTERNAL AC POWER(NOT SHOWN)

TWO OF EACH MOUNTING STUD SUPPLIED.

UNLESS SPECIFIED TOLERANCES	
DIMENSIONS IN INCHES	DIMENSIONS IN MILLIMETERS [IN BRACKETS]
DECIMALS XX ±.03 XXX ±.010 ANGLES ±2 DEGREES	DECIMALS X ±0.8 XX ±0.25 ANGLES ±2 DEGREES
FILLETS AND RADII .003 - .005	FILLETS AND RADII [0.07 - 0.13]

DRAWN	4/10/06	MFG	4/19/06
CHK'D	4/10/06	ENGR	4/10/06
APP'D	4/10/06	SALES	4/10/06
TITLE			
OUTLINE DRAWING MODEL 394A10 CALIBRATION SYSTEM			

PCB PIEZOTRONICS™	
3425 WALDEN AVE. DEPEW, NY 14043 (716) 684-0001 EMAIL: SALES@PCB.COM	
CODE IDENT. NO. 52681	DWG. NO. 3440
SCALE: FULL	SHEET 1 OF 1